

New England Biolabs Product Specification

Product Name: *Monarch[®] Genomic DNA Purification Kit*

Catalog #: *T3010S/L*

Kit Components: *Monarch[®] gDNA Tissue Lysis Buffer (T3011)*
Monarch[®] gDNA Cell Lysis Buffer (T3012)
Monarch[®] gDNA Blood Lysis Buffer (T3013)
Monarch[®] gDNA Binding Buffer (T3014)
Monarch[®] gDNA Wash Buffer (T3015)
Monarch[®] gDNA Elution Buffer (T3016)
Monarch[®] gDNA Purification Columns (T3017)
Proteinase K, Molecular Biology Grade (P8107)
Monarch[®] RNase A (T3018)
Monarch[®] Collection Tubes II (T2018)

Shelf Life: *18 months*

Storage Temp: *25°C*

Specification Version: *PS-T3010S/L v1.0*

Effective Date: *25 Apr 2019*

Assay Name/Specification (minimum release criteria)

Functional Testing (Blood Cell Lysis Buffer, Monarch[®]) - Genomic DNA is purified from 8 samples of whole pig blood using the Protocol for Extraction and Purification of Genomic DNA from Mammalian Whole Blood (non-nucleated), with $\geq 2 \mu\text{g}$ of gDNA being recovered in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.

Functional Testing (Cell Lysis Buffer, Monarch[®]) - Genomic DNA is purified from 8 individual aliquots of HeLa cells (10^6 cells/aliquot) using the Protocol for Extraction and Purification of Genomic DNA from Cultured Cells, resulting in $\geq 3 \mu\text{g}$ of gDNA being recovered in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.

Functional Testing (DNA Recovery and Purity) - Twenty-four Monarch[®] gDNA Purification Columns are tested with $5 \mu\text{g}$ of HindIII digested Lambda resulting in $\geq 75\%$ recovery in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.

Functional Testing (RNase A, Monarch[®]) - A $10 \mu\text{l}$ reaction in NEBuffer 4 containing 40 ng of fluorescein labeled RNA transcript and RNase A is incubated at 37°C . After incubation for 5 minutes, complete disappearance of the RNA substrate occurs at $\leq 1.0 \mu\text{g/ml}$ of RNase A, as determined by gel electrophoresis using fluorescent detection.



New England Biolabs Product Specification

Assay Name/Specification (minimum release criteria)
--

<p>Functional Testing (Tissue Lysis Buffer, Monarch®) - Genomic DNA is purified from 8 NEB10-Beta samples treated using the Supplemental Protocol for Genomic DNA Purification from Gram-negative Bacteria, resulting in $\geq 5 \mu\text{g}$ of gDNA being recovered in $\geq 80\%$ of the samples. OD 260/280 and 260/230 are ≥ 1.75 in $\geq 80\%$ of the samples.</p>

<p>* Individual Product Component Note - Standard Quality Control Tests are performed for each component included in Monarch® Genomic DNA Purification Kit and meet the designated specifications.</p>



Date 25 Apr 2019

Derek Robinson
Director of Quality Control

